

****Operating Temperature Limitations****

This page lists temperature limitations for various ASI Items

-MS2000 LCD: -20 to 60 C

From:

<https://asiimaging.com/docs/> - **Applied Scientific Instrumentation**

Permanent link:

https://asiimaging.com/docs/operating_temperature_limitations

Last update: **2021/09/23 17:15**

